

USER MANUAL

CALISO TEMPS

VOLUME TWO

BUILDER



Humpage Technology Ltd

46 Rockspray Grove
Walnut Tree

Milton Keynes MK7 7EA
United Kingdom

Tel: +44 (0) 1908 675472

Web: www.humpagetec.com

May 2015

CONTENTS

1.0	INTRODUCTION	1
2.0	DATA VIEWER	2
2.1	INTRODUCTION	2
2.2	VIEWING CALIBRATION DATA	2
2.3	THE RESULTS TAB	3
2.4	CALCULATING COEFFICIENTS	4
2.4.1	CALLENDAR VAN DUSEN	5
2.4.1.1	INPUTTING DATA INTO THE CVD	6
2.4.2	ITS-90	7
2.4.3	THERMOCOUPLES	8
2.4.3.1	SELECTING THERMOCOUPLES	8
2.4.3.2	PERFORMING THERMOCOUPLE CALCULATIONS	9
2.4.3.3	THERMOCOUPLE UNITS	9
2.4.4	REGRESSION	10
2.4.5	STANDARD COMPARISON	11
2.4.6	MULTI-POINT COMPARISON	12
2.4.7	GENERATION OF LOOK-UP TABLES	12
2.4.7.1	ITS 90 AND CALLENDAR VAN DUSEN	12
2.4.7.2	THERMOCOUPLE LOOK-UP TABLES	15
2.4.8	INSTRUMENT DATA TAB	16
2.5	SAVING EDITED CTD FILES	16
3.0	CREATING CERTIFICATE TEMPLATES	18
3.1	INTRODUCTION	18
3.2	MAKING CERTIFICATE TEMPLATES	18
3.2.1	STATIC LABEL TEXT	19

3.2.2	DATA FIELD TEXT	21
3.2.3	IMAGE COMPONENT	23
3.2.4	LINES AND BORDERS	24
3.2.5	MOVING AND RESIZING COMPONENTS	25
3.2.6	DELETING COMPONENTS	27
3.2.7	MULTI SELECT TOOL	27
3.2.8	PAGE SET UP	27
3.2.9	RE ORDERING PAGES	29
3.2.10	PRINT PREVIEW	30
3.2.11	VIEWING MULTI PAGE TEMPLATES	30
3.3	SAVING LOADING AND CREATING NEW TEMPLATES	31
3.4	LINKING CALIBRATION DATA TO A TEMPLATE	31
3.4.1	SELECTING DATA CONTAINERS	31
3.4.2	EXPORT BUTTON	33
4.0	MAKING CALIBRATION CERTIFICATES	34
4.1	INTRODUCTION	34
4.2	OPEN DATA FILE	34
4.3	LOAD TEMPLATE FILE	34
4.4	EXPORT DATA INTO TEMPLATE	34
4.5	PRINT CERTIFICATE	34
4.6	SAVING PERMANENT COPY OF CERTIFICATE TO FILE	36

SECTION 1

Introduction

Having acquired your valuable temperature calibration data using the Laboratory Interface you will now wish to perform calculations, analysis of the data, and produce calibration certificates and reports. These are the functions of Caliso Temps Builder software that will enable you to:

- View data acquired in previous calibration tests
- Perform data analysis and calculations, for example, calculation of Callendar Van Dusen coefficients
- Produce calibration certificates and reports including the generation of look-up tables for thermocouple corrections, and ITS-90 and Callendar Van Dusen resistance to temperature conversions, all based on the coefficients obtained from your calibration data

The first two are performed using Data Viewer, which enables you to import and process calibration data that you have acquired using the Caliso Temps Laboratory Interface. Calibration certificates are then produced from the calibration data via certificate templates, that you design for yourself, as will be explained later. You can create as many multi-page templates as you need to fit exactly the different types of instrument that you calibrate. For example, you can have a template for PRTs and a different template for thermocouples.

We have made every effort to make the complete process as simple and productive as possible, with help files and balloon tips to help guide you along the way.

Note:

If you have any comments or suggestions regarding Caliso Temps Builder we would be very happy to hear from you. You will find various ways to contact us in the main help menu.

As with all our software we very much hope that you enjoy using it.

Best Wishes

The Caliso Temps Team

SECTION 2 – The Data Viewer

2.1 Introduction

Some parts of the Data Viewer can also be found in Caliso Temps Laboratory Interface. In other words opening a Caliso Temperature Data file enables you to view all of the set-point run data that was acquired during your test. The Builder's Data Viewer, however, enables you to go a few steps further by giving you the ability to turn raw calibration data into calibration information of the following types:

- Callendar Van Dusen coefficients
- ITS-90 coefficients
- Thermocouple correction coefficients
- Polynomial regression coefficients

Calculated results are then saved in the Caliso Temperature Data file to provide a complete record of each device's calibration.

2.2 Viewing Calibration Data

Across the top of the screen you will see a row of buttons. Click the one titled "Open Data File". This will open up the standard Windows file open dialogue box, from where you select your data file.




Fig 2.2.1 Opening a Caliso Temperature Data File

Caliso Temperature Data files have the extension '.ctd'. The file name itself is generated automatically by Laboratory Interface in the following way:

(Date) – (Instrument Channel Name) + (Serial Port Name) (.ctd)

With reference to fig 2.2.1 we can see data files that were recorded on the 22nd of June, using an Isotech TTI-7 connected to COM8.

When your CTD file is opened you will see that the Run Data grid contains all the data for each of the test set-points. A row of tabs along the bottom of the grid allows you to view data for each set-point used in your calibration test.

Results	Run Data	Instrument Data		
	Channel Value	Channel Mean	Reference	Ref Mean
	156.2600000	156.260000	156.2600000	156.260000
	156.2600000	156.260000	156.2600000	156.260000
	156.2600000	156.254000	156.2600000	156.254000
	156.2600000	156.255000	156.2600000	156.255000
	156.2600000	156.254737	156.2600000	156.254737
	156.2500000	156.254444	156.2500000	156.254444
	156.2500000	156.254706	156.2500000	156.254706
	156.2500000	156.255000	156.2500000	156.255000
	156.2500000	156.255333	156.2500000	156.255333
	156.2500000	156.255714	156.2500000	156.255714
	156.2400000	156.256154	156.2400000	156.256154

Fig 2.2.2 Run Data

Note:

Individual Caliso Temperature Data files contain the run data for individual devices that were tested during the calibration run and the file name indicates which of the devices calibrated is now being viewed.

2.3 The Results Tab

When you click the results tab you will see on the left of the screen a grid, that looks rather like a spreadsheet, with three columns titled:

- Standard
- Unit under test
- Setpoint

The first column, 'Standard', shows, for each set-point, the average value of the reference temperature calculated for each set-point. Similarly the "Unit Under Test" column contains the average of the device's output: either in temperature units or probe units (mV for thermocouples or Ohms for PRTs). The third column, 'Setpoint', shows the nominal set-point value at which temperature bath or Block was set for that particular part of the run. The final column "uncertainty" enables you to enter a value for the measurement uncertainty of the data point. This can be included on your certificates.

Results	Run Data	Instrument Data	
Spread Tolerance 0.5 Offset Tolerance 0.5 Points for Mean 10			
Standard	Unit Under Test	Setpoint	Uncertainty
156.260000	156.260000	156.0000	
200.283333	200.283333	200.0000	
232.286667	232.286667	232.0000	
420.443333	420.443333	420.0000	
550.060000	550.060000	550.0000	
650.140000	650.140000	650.0000	
50.216667	50.216667	50.0000	

Fig 2.3.1 The Results Tab

The right hand side of the Run Data screen is the calculation panel. This is where you perform the analysis necessary to generate your calibration coefficients as will now be explained in some detail.

2.4 Calculating Coefficients

You have completed your calibration tests using Laboratory Interface and built a Certificate Template using the Certificate Designer. The next logical stage is to analyse this data and produce information, which is of direct relevance to your customers or users. In temperature calibration the data analysis usually produces sets of calibration coefficients, the exact nature of which is governed by the method in which raw sensor data is converted into temperature units. Caliso Temps Builder provides the means for you to calculate coefficients using the most widely used methods in temperature thermometry, which are:

IEC 751 for Industrial PRT's using the Callendar Van Dusen equation.

IT'S 90

- Thermocouple Correction Coefficients
- Regression coefficients for comparison calibrations
- Error calculations for temperature-temperature comparisons
- Multi channel comparison calibrations

The coefficients that are generated can be incorporated into your certificates as will be described later in this manual.




Fig 2.4 Selecting Calculation Type

To select which type of coefficient you will generate first of all select the Results tab and then click the Calculation Type button. You may then select from one of the four items, listed above, in the pop-up window that appears and then click OK.

Depending upon which calculation type you have selected you will see different data entry screens. Common to them all is the ability to drag-and-drop data from the Result Summary grid into the relevant data entry boxes or should you wish you may type this data in by hand. In all cases clicking the Calculate button will generate the appropriate coefficients.

2.4.1 Callendar Van Dusen

Callendar Van Dusen's equation is one of the most widely used methods of converting the resistance of an RTD (resistance temperature device) into temperature units (in this case $^{\circ}\text{C}$). Calibration points are normally as follows:

- A negative value (optional) and two positive temperatures.
- You will also be taking resistances at zero $^{\circ}\text{C}$. While only one reading would suffice adding more values will allow the mean to be used. If you have the water triple point value rather than 0°C simply enter 0.01 as the standard value and the corresponding resistance.

At least one nominal zero °C value is compulsory for us to make the calculation although if you do enter all three, in accordance with standard practice, the average of all three or how many you enter will be taken.

2.4.1.1 Inputting data into the CVD Data grid

As we have just said you can manually type the data into the grid although if you have recorded a CTD data file and opened it in the way described, you may drag-and-drop your selected entries into the CTD Data grid. By doing it this way it is possible to select your more favoured values to be used in the calibration.

	Standard	Unit Under Test
Nom 0°C(i)	0	101
Nom 0°C(ii)	0	102
Nom 0°C(iii)	0	103
Minus Value	-20	99
Plus (i)	50	101
Plus (ii)	75	102

Clear Swap Calculate

OK

Fig 2.4.1.1 Callendar Van Dusen data

Clicking the calculate button will produce the coefficients for A, B, C, and R0 bearing in mind that if no negative temperature has been entered the value of C will be set to zero.

A -0.00058823529411765

B 7.84313725490201E-6

C 0

R0 102

Nom 0 TMean = 0

Nom 0 RMean = 102

Fig 2.4.1.2 Calibration Coefficients

Hover the mouse over the data containers for the Callender Van Dusen type and you will see in a yellow hint box the container names that will appear on the temperature properties box for data field containers.


2.4.2 ITS-90

IT'S 90 is an involved method, the explanation of which is beyond the scope of this manual. There are several references available from sources such as that at www.isotech.co.uk. The basics of it are, however, that a PRT is calibrated using fixed-point cells. Each of the different types of fixed-point cell relies on the fact that the transitions, in other words freezing, melting or vapourisation, of carefully selected substances always occur at repeatable and accurately known temperatures. Cells are referred to using the name of the substance whose transition is being exploited, for example: aluminium zinc or tin. The most accurate forms of IT'S-90 calibration always use fixed-point cells although, for less accurate work, it may be acceptable to use temperature sources, such as furnaces or blocks, to reproduce the appropriate fixed-point cell.

0 to 231.928 (Tin)	Temperature °C	Resistance	R(H2Otp)	W	Wr
FP Indium 156.5985	156.5985				
FP Tin 231.928	231.928				
TP Mercury -38.8344	-38.8344				
TP Gallium 29.7646	29.7646				

Fig 2.4.2 ITS-90 Calculations

The coefficients calculated depend upon the range of the calibration. There is no need to remember these because Caliso Temps Builder will automatically set the correct coefficients depending upon the temperature range you select. In accordance with Standard IT'S 90 Practice, you can calculate two sets of coefficients relating to the high and low ranges stated in the definition. Also if required you may choose to define your nominal zero point as being either the water triple-point or the melting point of ice as shown in Figure 2.4.2.1 below.



Zero Reference

☒ Rtpw - Resistance at water triple point (0.010°C)
☐ Ro - Resistance at ice point of water (0.000°C)

Value (ohms)

☐ Use this value for all points

2.4.2.1 – Zero Reference Selection

As always it is possible to drag and drop the values of the Standard and Units Under test columns in the Result Summary grid on to your IT'S-90 Data grid. The temperature and resistance values will be copied across. The resistance at the triple point of water, however, will need to be entered manually.


Clicking the Calculate button will perform the necessary calculations to generate the coefficients required. Should you wish to clear the grid click the button marked clear. Also by selecting different temperature ranges you will clear the grid and if the fixed point cells box is checked the appropriate nominal temperature values (those of the fixed point cell) will be entered for you.

2.4.3 Thermocouples

Thermocouple calibration can take several subtly different forms but the most commonly used, and most useful, type is the voltage correction coefficient method. Basically what happens here is that, at the reference temperatures chosen, and for the correct type of thermocouple, the expected milli-volts are calculated for each temperature set-point and compared against the measured value to produce a milli-volt difference. A polynomial is then fitted through the curve generated from the reference temperature and difference values and a third order regression polynomial produced.

2.4.3.1 Selecting Thermocouples

First, click the Calculation Type button and select your calibration as thermocouple. Then, along the top of the thermocouple panel you will see a drop down list from which you select your device's thermocouple type.



Select Thermocouple Type		
	Ref Temp (°C)	Me
int 1	50.060000	0.0
int 2	60.080000	0.0
int 3	70.050000	0.0
int 4	80.050000	0.0
int 5	90.000000	0.0

Drop down list: B, D, E, J, K, N, R, S, T

Fig 2.4.3.1 Selecting thermocouple type

2.4.3.2 Performing Thermocouple Calculations

When you have selected your thermocouple type, you will now need to enter the reference temperatures and measured millivolts either manually by typing in or by dragging and dropping from the Results Summary grid. Clicking the Calculate button calculates a series of regression coefficients that are used to calculate the correction value for any reference temperature.

Calculation Type

Thermocouple: II Measured units: microVolts

	Ref Temp (°C)	Measured	Expected	Difference	Regressed
Point 1	0.180000	0.000009	4.667800	-4.7	-4.700
Point 2	29.850000	0.000792	788.970146	-789.0	-789.000
Point 3	125.050000	0.003526	3528.326004	-3528.3	-3528.300

Curve-fit: $Y = Cx^3 + Bx^2 + Ax + D$

☒ $Y = Cx^3 + Bx^2 + Ax$
☐ $Y = Bx^2 + Ax$

☐ Use regression for best-fit

Regression Order: Third

Coeff A: -26.1095036514159
 Coeff B: -0.00891893253054152
 Coeff C: -6.33283857225977E-5
 Coeff D: 0.0

Fig 2.4.3.2 Calculation of thermocouple data

Hover the mouse over the various data containers to familiarise yourself with them and remember that these are the names that will appear in the Properties box of any data field container used on a certificate template.

2.4.3.3 Thermocouple Units

Depending in the instrument that was used to measure the output voltage from the reference thermocouple, the value will be in either volts, milli-volts, or micro-volts. On the other hand, the NIST coefficients that are used to convert between temperature and voltage are always based on the micro-volt as the unit of measurement. It is important, therefore, that you specify the measured units so that the software can calculate accordingly.

Measured units		Volts	
Expected		milliVolts	Regressed
		Volts	
		microVolts	

Fig 2.4.3.3 Thermocouple Units

Select the appropriate units from the drop-down list near the top of the page (as shown in Figure 2.2.3.3).

Note:

The output coefficients are calculated on the basis of your chosen units. You will need to re-scale them accordingly if other units are required.

2.4.4 Regression

To select the regression calibration type click the calibration type button and select curve fit from the selected options and click the OK button (refer to fig. 2.4).

Regression Analysis is most commonly used for what are known as “comparison calibrations” where the temperature value indicated by the unit under test is compared to the same temperature indicated by a standard traceable unit. The coefficients generated here are regression coefficients that represent a best-fit polynomial between a standard unit and the unit under test over your range of set-point values.

As usual, you may either type these values in manually, or drag-and-drop them from the Results Summary grid. Click the Calculate button to generate your coefficients or the clear button to start again. If you have inadvertently entered the columns the wrong way round (i.e. you have entered the standard data in the units test column or vice versa) simply clicking the swap button will reverse these for you.

Curve-fit			
Data			
	Standard	Unit Under Test (x)	Regressed (y)
	50.060000	0.001330	50.0797398241
	60.080000	0.001607	60.0463904391
	70.050000	0.001889	70.0480662187
	80.050000	0.002175	80.0516190539
	90.000000	0.002464	90.0265588385
	99.930000	0.002755	99.9448018571
	109.860000	0.003048	109.81387644
	119.840000	0.003350	119.87367529
	129.790000	0.003650	129.764822943
	139.750000	0.003956	139.766

Fig.2.4.4 Regression analysis of calibration data

Regression Order: Third

Coeff A: 67810526.2534221

Coeff B: -1246310.40083788 Regression_A

Coeff C: 39201.1024063076

Coeff D: -0.0126614612669609

Clear Swap Calculate

Fig 2.4.4.1 Calculating regression data

2.4.5 Standard Comparison

One of the simplest forms of temperature calibration is to compare, over a range of temperatures, the value of the temperatures indicated by the unit under test compared to a standard reference unit at the same temperature. This is normally in the form of an error value.

[illegible]

Fig. 2.4.5 Standard Comparison

Adding data to the grid is done in the usual way i.e. dragging your chosen values from the Run Data Grid. Naturally you can include this data in your certificate.

2.4.6 Multi-point Comparison

There will be occasions when all of the channels recorded in your calibration test will need to be presented on a single calibration certificate, for example, a thermocouple array. This is easily done with Caliso Temps. With reference to Section 3.2.3 of the Interface manual, when specifying your test data files, if you selected the option “Create Multi-Channel Comparison File” option, the multi-channel summary will be available for you to include in your certificates.

Multi-Channel Comparison					
			Channel 1		Channel 2
	SetPoint	Reference	Value 1	Difference 1	Value 2
	25	24.95	-1000	-1024.950000	-1000
	26	26.225	-1000	-1026.225000	-1000
	27	27.18	-1000	-1027.180000	-1000
	28	28.2	-1000	-1028.200000	-1000

Fig. 2.4.6 Multi-point Comparison Data

Builder requires you to do very little to create a multi-point comparison certificate. Simply place the appropriate component on your template. In order to fit the page the data will be presented in blocks of four channels across. Should you wish to send CTD files and multi-point data to other people, make sure that you include the associated file having an MPF extension.

2.4.7 Generation of Look-Up Tables

2.4.7.1 ITS-90 and Callendar Van Dusen

A very useful feature of Caliso Temps is its ability to produce Look-Up tables for both platinum resistance thermometers and thermocouples. We will look at ITS90 and Callendar Van Dusen first as they are both resistance to temperature conversions. The first stage in the process is to produce your calculations based on your measured data and generate the appropriate set of coefficient which you then save for future reference. Having done this simply click the “Create Table” button as shown below and, in the case of ITS90, specify whether you wish to generate the table based on the high or low range coefficient.



2.4.7.1.1 Creating a Look-Up Table

Having clicked the button the table calculator window will appear which will contain the coefficients created. You will need to enter your latest value for the triple-point resistance of the PRT.


Note:

Another feature which is useful is to use the coefficient to calculate the temperature at any given resistance as you will also see in 2.4.7.1.2 below.

You also enter the serial number of the PRT (optional) and the range and increments you require. Having done this click the “Calculate” button.


2.4.7.1.2 Configuring the Look-Up Table

Look-Up tables are saved as “Tab-Delimited Text” files that can be viewed either in word processors or, more suitably, in spreadsheets such as Excel. You will be prompted to specify where you want to save this file.



2.4.7.1.3 Specifying where to save the Look-Up Table

Once a table has been created you will be asked if you would like to view the table in Excel. Click “Yes” if you do.



2.4.7.1.4 Viewing the Look-Up Table in Excel

You will then see the Look-Up Table as shown below. The first column shows °C and the subsequent columns are in increments as you specified.

Microsoft Excel - 23-45-4.txt

File Edit View Insert Format Tools Data Window Help

100%

Arial 10 B I U

N14 =

	A	B	C	D	E	F	G	H	I	J	K
1	Serial Number:	RTD 56-56-01									
2		0	0.1	0.2	0.3	0.4	0.5	0.6	0.7	0.8	0.9
3	10	-220.6601	-220.3947	-220.1301	-219.8663	-219.6033	-219.3411	-219.0796	-218.8188	-218.5587	-218.2993
4	11	-218.0406	-217.7826	-217.5252	-217.2684	-217.0122	-216.7566	-216.5016	-216.2471	-215.9932	-215.7398
5	12	-215.487	-215.2346	-214.9828	-214.7314	-214.4805	-214.2301	-213.9801	-213.7305	-213.4814	-213.2327
6	13	-212.9844	-212.7365	-212.489	-212.2418	-211.995	-211.7486	-211.5026	-211.2569	-211.0115	-210.7664
7	14	-210.5217	-210.2773	-210.0332	-209.7893	-209.5458	-209.3026	-209.0596	-208.8169	-208.5744	-208.3323
8	15	-208.0903	-207.8486	-207.6072	-207.366	-207.125	-206.8842	-206.6436	-206.4033	-206.1631	-205.9232
9	16	-205.6834	-205.4439	-205.2045	-204.9653	-204.7263	-204.4875	-204.2488	-204.0103	-203.7719	-203.5337
10	17	-203.2957	-203.0578	-202.82	-202.5824	-202.3449	-202.1076	-201.8704	-201.6333	-201.3963	-201.1594
11	18	-200.9227	-200.6861	-200.4496	-200.2131	-199.9768	-199.7406	-199.5045	-199.2685	-199.0325	-198.7967
12	19	-198.5609	-198.3253	-198.0897	-197.8542	-197.6187	-197.3833	-197.148	-196.9128	-196.6777	-196.4426
13	20	-196.2075	-195.9725	-195.7376	-195.5028	-195.2679	-195.0332	-194.7985	-194.5638	-194.3292	-194.0946
14	21	-193.8601	-193.6256	-193.3911	-193.1567	-192.9223	-192.688	-192.4536	-192.2193	-191.9851	-191.7508
15	22	-191.5166	-191.2825	-191.0483	-190.8141	-190.58	-190.3459	-190.1118	-189.8778	-189.6437	-189.4097
16	23	-189.1756	-188.9416	-188.7076	-188.4736	-188.2396	-188.0056	-187.7716	-187.5376	-187.3036	-187.0697
17	24	-186.8357	-186.6017	-186.3677	-186.1337	-185.8998	-185.6658	-185.4318	-185.1978	-184.9638	-184.7298

2.4.7.1.5 The Look-Up Table as it appears in Excel

2.4.7.2 Thermocouple Look-Up Tables

In a similar way to that described above you can generate tables for thermocouples. These can either be in the form of:

Temperature in °C vs thermocouple voltage

Thermocouple voltage correction vs thermocouple voltage

The Table is configured in exactly the same way as explained for PRTs and again as above you will be able to view your Look-Up Table in Excel.

Thermocouple Tables

Table Parameters

Serial Number: 12-34-44

Number Across = 10

Temperature Range

From: 10
To: 100

Increment

☐ 0.1
☒ 0.01
☐ 0.001

Output

☐ °C vs Output Voltage
☒ Voltage Corrections

2.4.7.2.1 Generating a Thermocouple Look-Up Table

2.4.8 The Instrument Data Tab

In Section 3.7 of the Interface Manual it was described how you can save key information relating to the units under test, the reference standard together with other information pertinent to the test. The Instrument Data Tab in Builder enables you to add to or modify this information.

The screenshot shows the 'Instrument Data' tab in the Builder software. It is divided into four main sections:

- Standard:** Contains fields for Serial Number (Std PRT 12-32-12), Coeff A (0), Coeff B (0), Coeff C (1), Coeff D (0), ITS-90 WAI, and ITS-90 Wtp.
- User Comments:** Contains six Test Comments fields (1 through 6).
- Unit Under Test:** Contains fields for Type (dropdown), Serial Number (TC II 456-1), Channel Name (TTI-7 B4 COM8), Model, Job Reference, and Comment.
- Test:** Contains fields for Test Date (17/01/2005), Test Time (17:27:04), and a Test Memo text area.

2.4.8. The Instrument Data Tab

All of the information can be included in your certificate templates by selecting appropriate data fields on your template. For those performing calibrations for other people you can save a job reference number for each of the devices.

2.5 Saving edited CTD files

The CTD data file that you created using the data-logger program contains, among other things, the following data:

Calibration data

Information on the units-under test and reference instrument

Clearly, because at the time the calculations described above had not been carried out, values of coefficients etc are blank. Once you have performed the necessary calculations, you will want to save the extended data so that you don't have to keep repeating the work you have just done.

To save your updated CTD file you can use the 'Save Data File' menu items as shown in fig 2.5 below.



S
S
S




Fig 2.5 Saving edited CTD data

3.0 CREATING CERTIFICATE TEMPLATES

3.1 Introduction

Caliso Temps Builder enables you to design and save calibration certificate templates in a WYSIWYG* environment.

* What You See Is What You Get

Certificate templates are constructed using one of several types of component:

- Static label text component
- Data Field component
- Image component
- Graphic component. (for drawing lines, boxes, and borders)

These components can be moved around the page, either singly or in groups, simply by dragging to the desired position. Each component is described later.

Templates can be saved for future use, and there is no limit to the number of templates you can design and save.

3.2 Making Certificate Templates

Whilst some sample templates are included we feel sure that you will enjoy making calibration certificates using Caliso Temps Builder. All the tools needed to produce great looking certificates quickly and easily are here. Your certificate templates can contain several pages – you could, for example, have some pages in portrait and others in landscape, incorporate logos, text and, of course, calibration data. The quickest way to learn how to make certificates is to use the software, so that's what we'll do.

To start, select 'File' and then 'New Template' from the main program menu. What you see is a preview of an, as yet, blank template.




Fig 3.2 The Certificate Designer

Click the 'Edit' button to open the 'Certificate Designer' window.

What you see are:

The template page: scaled to the size of your printer's default paper size. You can also see that the page has a 1cm graticule to aid component alignment, a vertical ruler and horizontal ruler both with 1cm graduations.

The Properties Box and Component Palette: where you edit the properties of your template components (more of this later).

3.2.1 Static Label Text

When you design a calibration certificate, you will no doubt wish to give it a title caption, add a few comments, and to put labels next to the data fields (e.g. "Device Type", "Last Calibration Date", etc). This text, which does not change from device to device, is referred to as Static Label Text.

Adding a New Label Component.

Select the Label component from the toolbar as shown below.

When you have selected the Label component, click the certificate page in the place where you wish the label to appear.




Fig 3.2.1a Adding a new label

Label components have property values that you can set. These are:

- Text The words that appear in the label
- Font Typeface, size, bold italic colour etc.
- Alignment Left justify, right justify, centre

These are changed using the Label Properties toolbar.

If a number of label components are selected simultaneously using the multi-select option, all selected components will be given the new properties. Furthermore, using the multi-select option a number of components can be resized simultaneously, or moved as a group to a different part of the page.

Once added to a certificate template, a label's properties can be edited at any time, by clicking on the label you wish to alter, and then entering new values.




Fig 3.2.1b An example of a static label

Copying and Pasting Text into a Label

The static text component can, should you wish, hold a large amount of text. If this already exists in another document, there is no need to re-type it, just use the copy/paste function as follows:

Copy your text from your source document onto the Clipboard (usually Edit – Copy)

Right-click your text label and select 'Paste'. The text will now appear in the label.




Fig 3.2.1c Pasting text into a static label

A Note on Word Wrapping

Word wrapping is the process where by any text too wide to fit onto a single line, will automatically be placed on several lines. This is particularly useful, for when creating a label that contains an address, because the entire entry can be made using a single component. Static Text Labels have this word wrapping capability. Do be careful when using it, however, and, if in doubt, use one component for a single line of text only.

3.2.2 Data Field Text

This component links the calibration certificate template to the data in the Caliso Temperature Data file that you currently have open.

With reference to Section 2, where the Data Viewer was discussed, controls such as the Run Data grid, which contain information stored in the file, are called 'Data Containers'. The data displayed in the Data Field will depend on which Data Container is selected.

There are many Data Containers in Data Viewer. Some contain blocks of data, while others contain single pieces of information; here are a few examples:

- Run Data
- CVD_Data
- CVD_A
- CVD_R0
- Instrument Serial Number
- Date

When a template is used to generate certificates for different devices, this component will contain data from the currently open Caliso Temperature Data file.

Note:

The certificate is only updated with data from the current file when the 'Export' button is clicked. For more information on this, see Section 3.4.2 later.

Adding a New Data Field Component

1. Select the Data Field component from the toolbar as shown below.
2. When you have selected the Data Field component, click the certificate page in the place where you wish the data to appear.




Fig 3.2.2a Adding a new Data Field component

3.2.3 Image Component

Image components allow you to place bitmap or JPEG images onto your calibration certificate. Most commonly, this will be your company logo, or the approval mark of a notified body.

In accordance with the wishes of most trademarks, the images are not resizable. This because trademarks usually have rules for their use, which will include size, colour and aspect ratio.

To add a bitmap or JPEG to a certificate, firstly, select the Image component from the toolbar.




Fig 3.2.3a Adding an image

Then click the place on the page where you wish the image to appear (don't worry about being too accurate, because you can always move it later).

A dialog box will then appear, enabling you to select your required file.

Resizing the component will have no effect on the size of the printed image.

Note

To stop template files becoming too large, the files do not contain the bitmaps, or JPEGs, themselves, but the file path to the file. This means that if the file is not in the place specified in the template, the image cannot be drawn.

Please bear this in mind when relocating files, or sending templates to others. You are advised to keep certificate bitmaps and JPEGs in a single folder accessible to all users.

Once saved as a Windows metafile, the certificate image becomes a permanent record.

3.2.4 Lines and Borders

A few embellishments, such as lines, boxes and borders, can greatly improve the appearance of any document. Calibration certificates are no different.

Caliso Temps Builder has a Graphic component to do this for you.

Adding a Graphic Component

Select the Graphic component from the toolbar as shown below.

When you have selected the Graphic component, click the certificate page in the place where you wish the label to appear (don't worry about being too accurate, because you can always move it later).




Fig 3.2.4a Adding a Graphic Component

A rectangular box of default dimensions will now appear in the place where you clicked the page. You then use the Properties window to turn this, default, box into a line, box or border of your own making.

Graphic Properties

The Properties box allows you to create a wide range of graphics:

Style	Select from one of the following options
Box	A rectangle with square corners
Rounded Box	A rectangle with rounded corners
Ellipse	Elliptical box, note that a circle is a special case of an ellipse, its box is square
Vertical Line	
Horizontal Line	
Position (%)	Specify the size and location of your Graphic. All values are percentage dimensions of the actual paper size.

Line Width

Select the required thickness in pixels. You can also specify whether your graphic is made up of single or double lines.

Colour

Specify the line colour for your Graphics.




Fig 3.2.4b Graphic Properties

3.2.5 Moving and Resizing Components

You will no doubt wish, at some time, to relocate or resize the components on your certificate (but don't forget the comments on images and trademarks made earlier).

Components, with the exception of Graphic components, can be moved or resized if they have focus. This is explained in the pictures below:



3.2.5a Component WITH focus



3.2.5b Component WITHOUT focus

To resize a component

Place the mouse cursor over one of the focus tabs, and drag the outline of the component to the required size.

To move a component

Place the mouse cursor, over the control, but not over the tabs, and drag the component to the required location.

Moving Several Components at Once

It is possible to move several components at once to a different position on the certificate, whilst maintaining their positions relative to each another. This can be done in several ways using the multi-select option.

You may also align the left-hand sides of a number of components to form a data column by clicking the align left sides button on the Multi-Selected Control Options properties window. For more details, see the topic entitled "The Multi-Select Tool".




Fig 3.2.5c Multi-component selection


To enable multiple-moving, click the Multi-Select button. Then click all of the components you wish to move. Then drag any one of the focussed components to where you wish to be moved. When you release the mouse button, the selected components will relocate themselves.

Alternatively, you may move the mouse cursor to anywhere on the page except on a Label, Data Field, or Image component, and then hold down the shift key and drag the components to their new position. You will notice that the mouse cursor changes to a hand when you press the mouse button.

Since Graphics will usually be page borders, these must be moved individually, rather than as part of a multi-selection.

3.2.6 Deleting Components

The Delete button, shown below, enables you to delete the focussed component from a certificate. You will be prompted to confirm this action before it is carried out.



3.2.6a Deleting the selected component.

Note:

The Delete button is disabled whenever the Multi-Select button is IN.

3.2.7 The Multi-Select Tool

There will be times when you will want a number of labels or Data Fields to be the same size, or have the same font, or to be aligned to form a column (especially if you are adding calibration measurements to your certificate).

This is best-done using the Multi-Select Tool. (See fig 3.3.5c)

The multi-select tool allows you to perform simultaneously a range of operations on selected labels and Data Fields.

To illustrate by example, if we wish to set a number of labels' widths to 103, and to align them into a column:

- Click the multi-select button.
- Select the required components by clicking.
- Enter 103 into the width box.
- Click the Same Width button.
- Enter the new left co-ordinate into the left box
- Finally, click the Align Left Sides button.

3.2.8 Page Set-up

A standard Widows Print Page set-up dialogue can be accesses by selecting 'Page Setup' from the 'File' main menu.

From here you can:

- Select your printer (if you have more than 1 connected or are running on a network)
- Select paper orientation
- Select paper size




Fig 3.2.8a Page Setup Menu Item

There are 2 important points to note:

The selected paper orientation will apply to the current page; in other words, you can select each page's orientation individually.

The paper size selected will apply globally; i.e. the selected paper size will apply to the whole template.




Fig 3.2.8b The Print Setup Dialogue

The selected printer is not saved within the template file, but the paper size and orientation is.

3.2.9 Re-ordering Pages

It will not always be the case that the order in which you designed the pages of a multi-page certificate template is the order in which you want them to be printed. Pages can be re-arranged at any time and the revised page order saved as part of the template. To do this, select 'File' and then 'Re-Order Pages' from the main menu. The Page Re-Ordering window will then appear.




Fig 3.2.9a Re-Ordering Template Pages

To re-order a page, select it in the left-hand list box and then use the Up and Down buttons to reposition it in the right-hand list box (see fig 3.2.9b). Click OK to implement your changes, or Cancel to revert to the original order.




Fig 3.2.9b Re-Ordering Pages of a Template.

You will also notice that the right-hand side of the window contains a preview of the selected template page.

3.2.10 Print Preview

A preview of how the printed certificate will appear can be obtained at any time by clicking the Print Preview button. The preview will show either the selected Data Container or the last exported data and is not intended to be a calibration certificate for any particular device.




Fig 3.2.10 Obtaining a print preview.

When you click the Print Preview button the preview window appears.

In addition to an image of your certificate, you will also see several radio-buttons, which enables you to zoom to the preview to your required size.

3.2.11 Viewing Multi-Page Templates

Some of your templates will consist of more than one page.

When a multi-page template file is open, you can navigate around the various pages by selecting the page from the 'Window main menu item



Fig 3.2.11 Selecting Template Pages

The current window is indicated by a tick mark as shown.

3.3 Saving Loading and Creating New Templates

Creating new templates, loading templates you might have designed earlier or saving your new template are all done through the main Certificate Builder menu. The options are presented when you click the File menu item.




Fig 3.3 Saving, loading and creating templates.

- New Clears the old template ready to add new components
- Open Loads an existing template
- Save Saves template overwriting old version
- Save As Saves template to new file leaving old copy unchanged

There is also a menu item that enables you to load your most recently used templates without having to browse for them.

3.4 Linking Calibration Data to a Template

Because this part of the program is the key to building successful and accurate calibration certificates, it is sufficiently important to justify a section of it's own. This is because it is how real calibration data is placed on to your calibration certificates. To illustrate how to achieve this, follow the next few steps:

3.4.1 Selecting Data Containers

Step 1

Click the button at the top of the screen marked 'Open Data File' and, using the Windows file dialogue box, select one of your Caliso Temperature Data files. Your data will then appear as described earlier in this manual (see fig 2.2.1).

Note:

If you have not yet created any CTD files of your own, then use some of the sample files provided.

Step 2

Click the 'Run Data' tab and hover the mouse over the Run Data grid. A yellow hint box will appear that contains the words 'Run_Data'. This is the identity of the Data Container to which data will be linked in your template and, hence, certificate (see fig 3.2.2c).

Step 3

Now click 'File' and then 'New Template' from the main program menu as shown in fig 3.4.1. Then click the 'Edit' button to show the Certificate Designer.

Place a new Data Field on to your blank template page, and then resize it so it fills approximately the top half of your page (this will enable you to see some of the block-data contained in grid type containers).




Fig 3.4.1 Opening a new certificate template

In the Properties Box you will see a list of the available Data Containers. Scroll down the list and select the "Run_Data" entry (see fig 3.2.2a). This has linked the Run Data grid to the Data Field component you have just added your certificate template. Now, whenever you use this template to build a certificate, the contents of the Run Data grid will be placed on your certificate, all you have to do is open the correct CTD file and 'export' the data.

To demonstrate this, click the 'Export' button. You will now see the run data in the box as shown fig 3.4.2.

Note:

Please refer to the section that talks about word wrapping and also note that it may be well to select a small font to display as much data as you require. Always ensure that your control is wide enough to fit the columns, or the tabular format will be lost.




Fig 3.4.2 A Data Field component containing Run Data table

Now try selecting other Data Containers, e.g. CVD_data, from the Data Field properties box. Click the 'Export' button each time and you will see that the component now contains data from the other selectable fields. Try this for as many components as are available to get the general feel of what is happening.

3.4.2. The Export Button

As you have seen, the function of the 'Export' button is to transfer data from the Data Viewer into the selected Data Containers on your template. You must do this each and every time you select a new CTD file.

SECTION 4 Making Calibration Certificates

4.1. Introduction

Congratulations!

You are now about to take the final, and in some ways, simplest step. You have used Caliso Temps to set-up an automated temperature calibration, and saved the data in CTD format. You then used Caliso Temps Builder to design a certificate template that exactly matches your requirements, with Data Containers, text and graphics. All that remains to do now is to use all of this to make calibration certificates.

Most of the steps necessary have been described earlier, so, to avoid repetition, the relevant section numbers will be given where this is the case.

4.2 Open Data File

See section 2.2.

4.3 Load Template File

See section 3.3.

There will seldom be any need to see the Certificate Designer once you are happy with your template layout. So, when all you are doing is making certificates from a template, it will not be visible. Should you want to make a few changes, just click the 'Edit' button.


4.4 Export Data into your Template

See sections 3.4.1 and 3.4.2

4.5 Print Certificate

To obtain a printed copy of your finished certificate, select 'File' and then 'Print' from the main program menu or click the 'Print' button located at the top of the screen.

Fig 4.5 Printing Certificates



Note:

Print Order.

Let us say, for example, that you have a 5 page certificate, and specified the page order 1, 2, 3, 4, 5, i.e. 1 at the top and 5 at the bottom. You will see that the pages are printed in the order 5, 4, 3, 2, 1, thus presenting the pages in the correct order without the need for re-shuffling.

4.6 Saving a Permanent Copy of your Certificates to File

In addition to keeping paper copies of your certificates, you may also want to maintain copies as computer files. Builder can do this for you. Each page of your certificate can be saved in standard Windows Metafile format. This means that you view images at any time in applications such as MSWord, or PaintShop Pro. You could also e-mail them as an attachment.

To save your certificates, click 'File' and then 'Save AS WMF' on the main program menu, or click the 'Save As WMF' button located near the top of the screen.




Fig 4.6.1 Saving Certificate Images

The standard Windows Save dialogue box will then appear, from where you enter your file name in the usual way. Builder will then automatically adjust your chosen name to indicate the page number of a multi-page document in the following way: Imagine that you have selected to save a 4-page certificate as:

c:\My Documents\MyCertificate.wmf

Builder will modify this to create 4 files whose names incorporate the correct page number as follows:

c:\My Documents\MyCertificate-1.wmf

c:\My Documents\MyCertificate-2.wmf

c:\My Documents\MyCertificate-3.wmf

c:\My Documents\MyCertificate-4.wmf




Fig 4.6.2 Selecting Picture File Names

A note of caution

Certain applications, MS Word being one of them, enable you to edit and modify Windows Metafiles. If you will be saving your certificate files on a network drive, then you might wish to consider allowing read-only access to prevent unauthorised tampering.