

Renewal of HIOKI's world-leading battery tester



BT3563A



The de facto standard for accurate measurement of large battery packs up to 300V



BT3562A



The de facto standard for accurate measurement of large xEV and ESS battery cells, as well as 72V and 96V packs



BT3561A



For high-performance battery cell applications in electric bikes and power tools. Ideal for battery packs of up to 60V

Designing automatic battery testing systems is easier and faster than ever before

- Double the total line resistance, so measurement errors are less likely to occur when using long measurement cables
- Stable operation regardless of increased total line resistance due to probe and relay degradation
- LAN is equipped as a standard for easy system design and layout, and excellent noise resistance for stable operation
- Improved electrostatic resistance as a countermeasure against electrostatic charges during battery transport on a production line

Lineup

Application		Acceptance inspection of general-purpose, small cells installed in a high-speed sorters	Fully automated production line testing of small cells for power motors or small packs of up to 60 V	Fully automated production line testing of large cells for xEVs or mid-size packups of to 100 V	Fully automated production line testing of large packs for xEVs or large packs up of to 300 V
Model		3561, 3561-01	BT3561A	BT3562A	BT3563A
Appearance					
Measurement method		AC four-terminal method	AC four-terminal method	AC four-terminal method	AC four-terminal method
Measurement frequency		1 kHz ± 0.2 Hz	1 kHz ± 0.2 Hz	1 kHz ± 0.2 Hz	1 kHz ± 0.2 Hz
Measurement parameters	Resistance measurement ranges	3 m Ω	N/A	N/A	3.1000 m Ω , 0.1 $\mu\Omega$, 100 mA
		30 m Ω	N/A	31.000 m Ω , 1 $\mu\Omega$, 100 mA	31.000 m Ω , 1 $\mu\Omega$, 100 mA
		300 m Ω	310.00 m Ω , 10 $\mu\Omega$, 10 mA	310.00 m Ω , 10 $\mu\Omega$, 10 mA	310.00 m Ω , 10 $\mu\Omega$, 10 mA
		3 Ω	3.1000 Ω , 100 $\mu\Omega$, 1 mA	3.1000 Ω , 100 $\mu\Omega$, 1 mA	3.1000 Ω , 100 $\mu\Omega$, 1 mA
		30 Ω	N/A	31.000 Ω , 1 m Ω , 100 $\mu\Omega$	31.000 Ω , 1 m Ω , 100 $\mu\Omega$
		300 Ω	N/A	310.00 Ω , 10 m Ω , 10 $\mu\Omega$	310.00 Ω , 10 m Ω , 10 $\mu\Omega$
	Max. display, resolution, measurement current	3 k Ω	N/A	3.1000 k Ω , 100 m Ω , 10 $\mu\Omega$	3.1000 k Ω , 100 m Ω , 10 $\mu\Omega$
		Basic accuracy	3 m Ω range 30 m Ω range or more	N/A $\pm 0.5\%$ rdg. ± 5 dgt.	$\pm 0.5\%$ rdg. ± 10 dgt. $\pm 0.5\%$ rdg. ± 5 dgt.
	Voltage measurement ranges	6 V	N/A	6.00000 V, 10 μV	6.00000 V, 10 μV
		20 V	19.9999 V, 100 μV	N/A	N/A
		60 V	N/A	60.0000 V, 100 μV	60.0000 V, 100 μV
		100 V	N/A	100.000 V, 1 mV	N/A
	Max. display, resolution	300 V	N/A	N/A	300.000 V, 1 mV
		1000 V	N/A	N/A	N/A
	Basic accuracy	$\pm 0.01\%$ rdg. ± 3 dgt. ^{*1}	$\pm 0.01\%$ rdg. ± 3 dgt.	$\pm 0.01\%$ rdg. ± 3 dgt.	$\pm 0.01\%$ rdg. ± 3 dgt.
Response time ^{*2}		3 ms	10 ms	10 ms	10 ms
Sampling period ^{*3}		Ω or V	4 ms, 12 ms, 35 ms, 150 ms	4 ms, 12 ms, 35 ms, 150 ms	4 ms, 12 ms, 35 ms, 150 ms
EX.FAST, FAST, MEDIUM, SLOW		QV	7 ms, 23 ms, 69 ms, 252 ms	8 ms, 24 ms, 70 ms, 253 ms	8 ms, 24 ms, 70 ms, 253 ms
Allowable total line resistance ^{*2,4} (within accuracy)		SENSE line	N/A, N/A, 20 Ω , 20 Ω	N/A, 4 Ω , 30 Ω , 30 Ω	4 Ω , 4 Ω , 30 Ω , 30 Ω
Ranges: 3 m Ω , 30 m Ω , 300 m Ω , 3 Ω		SOURCE line	N/A, N/A, 20 Ω , 20 Ω	N/A, 4 Ω , 20 Ω , 40 Ω	4 Ω , 4 Ω , 20 Ω , 40 Ω
Allowable total line resistance ^{*2,4} (error detection)		SENSE line	N/A, N/A, 20 Ω , 20 Ω	N/A, 6 Ω , 30 Ω , 30 Ω	6 Ω , 6 Ω , 30 Ω , 30 Ω
Ranges: 3 m Ω , 30 m Ω , 300 m Ω , 3 Ω		SOURCE line	N/A, N/A, 20 Ω , 20 Ω	N/A, 6 Ω , 20 Ω , 200 Ω	6 Ω , 6 Ω , 20 Ω , 200 Ω
Open terminal voltage					
Ranges: 30 m Ω or less, 300 m Ω , 3 Ω or more		N/A, 7 V, 7 V peak	25 V, 7 V, 4 V peak	25 V, 7 V, 4 V peak	25 V, 7 V, 4 V peak
Interface	LAN (TCP/IP, 10BASE-T/100BASE-TX)	N/A	✓	✓	✓
	RS-232C ^{*5} (Max. 38.4 kbps)	✓ (9.6 kbps fixed)	✓	✓	✓
	USB	N/A	N/A	N/A	N/A
	GP-IB	✓ (3561-01 Only)	N/A	N/A	N/A
	EXT. I/O (37-pin Handler interface)	✓	✓	✓	✓
Function	Analog output (DC 0 V to 3.1 V)	N/A	✓	✓	✓
	Contact check	✓	✓	✓	✓
	Zero adjustment (± 1000 counts)	✓	✓	✓	✓
	Pulse measurement	✓	✓	✓	✓
	Comparator	Hi/ IN/ Lo	Hi/ IN/ Lo	Hi/ IN/ Lo	Hi/ IN/ Lo
	Statistical calculations	Max. 30,000	Max. 30,000	Max. 30,000	Max. 30,000
	Delay	✓	✓	✓	✓
	Average	2 to 16 times	2 to 16 times	2 to 16 times	2 to 16 times
	Panel saving/loading	126	126	126	126
	Memory storage	400	400	400	400
	LabVIEW® driver ^{*6}	N/A	✓	✓	✓
Applicable standards		Safety: EN61010 EMC: EN61326 Class A	Safety: EN61010 EMC: EN61326 Class A	Safety: EN61010 EMC: EN61326 Class A	Safety: EN61010 EMC: EN61326 Class A
Effect of radiated radio-frequency electromagnetic field (10 V/m) ^{*7}		Resistant	Resistant	Resistant	Resistant
Effect of conducted radio-frequency electromagnetic field					
10 V		N/A	Resistant	Resistant	Resistant
0.15 MHz to 80 MHz, 80% AM		3 V	Resistant	Resistant	Resistant
CE		✓	✓	✓	✓
CSA ^{*8}		N/A	Certification in progress	Certification in progress	Certification in progress

*1: rdg. stands for *reading*, dgt. stands for *digits* *2: Typical value *3: When the power supply frequency is 60 Hz

*4: Total line resistance = wiring resistance + contact resistance + DUT resistance *5: Available as printer I/F

*6: LabVIEW® Driver is a registered trademark of National Instruments Corporation *7: Test conditions were 80 MHz to 1 GHz at 10 V/m and 1 GHz to 6 GHz at 3 V/m, all at 80% AM

*8: Canadian Standards Association

Note: Company names and product names appearing in this brochure are trademarks or registered trademarks of various companies.

HIOKI
HIOKI E. E. CORPORATION

HEADQUARTERS

81 Koizumi,
Ueda, Nagano 386-1192 Japan
<https://www.hioki.com/>



Scan for all
regional contact
information

DISTRIBUTED BY

CALPLUS
Die Kompetenz in der Messtechnik

CalPlus GmbH - Zentrale
Heerstraße 32
14052 Berlin
Tel.: 030 214982-0
Fax: 030 214982-50
office@calplus.de
www.calplus.de

CalPlus GmbH
Normannenweg 30
20537 Hamburg
Tel.: 040 3039595-0
Fax: 040 3039595-50
scopeshop@calplus.de
www.calplus.de