

A40B Precision Current Shunts Adapters, cables and connectors

Technical Data

Commonly used leads and adapters

A40B-LEAD/4MM

N to 4 mm double banana connector lead



This lead permits connecting the A40B shunt's N-type connectors to instrumentation that uses dual banana plug connection terminals. Using two leads permits connecting the shunt's voltage output to a voltage measuring instrument and also connecting the shunt's current input to a UUT current source rated 20 A or less.

A40B-ADAPT/LCN

LC female to N male inter-series adapter



This adapter enables connecting the N-type input connectors of A40B shunts rated 20 A or less with sources that have male LC current connectors.

A40B-LEAD/N

N male to N male lead



This lead permits connecting the A40B shunt's N-type output connector to a voltage measurement device with a female N-type input connector.

A40B-ADAPT/LC

LC male to LC male adapter



This adapter permits connecting current sources with LC female current connectors to A40B LC female shunt input connectors (the 50 A and 100 A shunts). This can also be used with the A40B ADAPT/LCN if a female mating connector is required.

Recommended lead and adapter configurations for the A40B shunt input (current) connection

Output connector of the source	Input connector of the shunt (shunt rating)	
	N-type (Shunts – 20 A or less)	LC-type female (50 A or 100 A Shunts)
Dual banana plug	A40B-LEAD/4MM lead	N/A
LC-type male	A40B-ADAPT/LCN adapter plus a user-supplied cable with a female LC connector	User-supplied cable with a female LC connector
LC-type female	A40B-ADAPT/LC adapter plus a user-supplied cable with a male LC connector	A40B-ADAPT/LC adapter plus a user-supplied cable with a male LC connector
N-type	A40B-LEAD/N lead	A40B-ADAPT/LC adapter plus a user-supplied cable with a male LC connector

Recommended lead and adapter configurations for the shunt output (voltage measurement) connection

Input connector of the measurement instrument	Output connector of the shunt (N-type female)
Dual banana plug	A40B-LEAD/4MM lead
N-type	A40B-LEAD/N N lead

Specialized accessories

(Recommended When Calibrating A40B Shunts)

A40B-CAL/LC
High current adapter



The high current adapter is used when calibrating the A40B shunts at current levels of more than 2 A. It connects together the current source and two shunts for calibration measurements. This adapter uses LC connectors. In this configuration, a certified or reference shunt is connected in series with a shunt that is being calibrated, which in turn are in series with a current source.

A40B-CAL/N
Low current adapter



The low current adapter is used to calibrate shunts at currents levels of 2 A or less. It connects together the current source and two shunts for calibration measurements. This adapter mates with N-type connectors to the shunts and a dual banana connection to the source. In this configuration, a certified or reference shunt is connected in series with a shunt that is being calibrated, which in turn are in series with a current source.

Fluke. Keeping your world up and running.®

Fluke Corporation
PO Box 9090, Everett, WA 98206 U.S.A.

Fluke Europe B.V.
PO Box 1186, 5602 BD
Eindhoven, The Netherlands

For more information call:
In the U.S.A. (800) 443-5853 or
Fax (425) 446-5116
In Europe/M-East/Africa +31 (0) 40 2675 200 or
Fax +31 (0) 40 2675 222
In Canada (800)-36-FLUKE or
Fax (905) 890-6866
From other countries +1 (425) 446-5500 or
Fax +1 (425) 446-5116
Web access: <http://www.fluke.com>

Modification of this document is not permitted without written permission from Fluke Corporation.

©2009 Fluke Corporation.
Specifications subject to change without notice.
Printed in U.S.A. 1/2009 3439147 D-EN-N Rev A
Pub_ID: 11519-eng, Rev 01